
ADDENDUM 002
REQUEST FOR BID
Dynamic Shear Rheometer
RFB # 5-160923LT

Bidders should acknowledge receipt of Addendum 002 (TWO) by signing and including it with the original bid. The due date for receipt of bids ***is NOT changed by this Addendum and remains at 2:00 p.m., Central Time, September 23, 2016***. Accordingly, the following clarification, and/or additional information, are believed to be of general interest to all potential bidders. All other terms and conditions remain unchanged and in full force.

Name and Title of Signer (Print or type)	Name and Title of Department Authority Lori Tackett General Services Specialist
Bidder/Offeror Signature _____ (Signature of person authorized to sign)	Department of Transportation <i>Lori Tackett</i> _____ (Authorizing Signature)
Date Signed:	Date Signed: September 22, 2016

Question: Regarding the section "ADDITIONAL REQUIREMENTS" on page 10 of 12, can you please clarify/confirm that "During the two year warranty period semi-annual calibrations and service will be included" means that we must provide calibrations at month 6, month 12, month 18, and month, 24, for a total of four calibration service visits required?

Answer: We expect calibrations at 6, 12, 18, and 24 months to be included in the bid. All of these would fall during the two year warranty period.

